


AG5 Refractometer Calibration Fluid Refractive Index Temperature Relationship

Temperature (°C)	0.0	0.1	0.2	0.3	0.4	0.5	0.6	0.7	0.8	0.9
10	1.34114	1.34113	1.34112	1.34111	1.34111	1.34110	1.34109	1.34108	1.34108	1.34107
11	1.34106	1.34105	1.34105	1.34104	1.34103	1.34102	1.34101	1.34101	1.34100	1.34099
12	1.34098	1.34097	1.34097	1.34096	1.34095	1.34094	1.34093	1.34093	1.34092	1.34091
13	1.34090	1.34089	1.34089	1.34088	1.34087	1.34086	1.34085	1.34084	1.34084	1.34083
14	1.34082	1.34081	1.34080	1.34079	1.34078	1.34078	1.34077	1.34076	1.34075	1.34074
15	1.34073	1.34072	1.34072	1.34071	1.34070	1.34069	1.34068	1.34067	1.34066	1.34065
16	1.34064	1.34064	1.34063	1.34062	1.34061	1.34060	1.34059	1.34058	1.34057	1.34056
17	1.34055	1.34054	1.34054	1.34053	1.34052	1.34051	1.34050	1.34049	1.34048	1.34047
18	1.34046	1.34045	1.34044	1.34043	1.34042	1.34041	1.34040	1.34039	1.34038	1.34037
19	1.34036	1.34035	1.34034	1.34033	1.34032	1.34031	1.34030	1.34029	1.34028	1.34027
20	1.34026	1.34025	1.34024	1.34023	1.34022	1.34021	1.34020	1.34019	1.34018	1.34017
21	1.34016	1.34015	1.34014	1.34013	1.34012	1.34011	1.34010	1.34009	1.34008	1.34007
22	1.34006	1.34005	1.34004	1.34003	1.34001	1.34000	1.33999	1.33998	1.33997	1.33996
23	1.33995	1.33994	1.33993	1.33992	1.33991	1.33989	1.33988	1.33987	1.33986	1.33985
24	1.33984	1.33983	1.33982	1.33980	1.33979	1.33978	1.33977	1.33976	1.33975	1.33974
25	1.33972	1.33971	1.33970	1.33969	1.33968	1.33967	1.33965	1.33964	1.33963	1.33962
26	1.33961	1.33960	1.33958	1.33957	1.33956	1.33955	1.33954	1.33952	1.33951	1.33950
27	1.33949	1.33947	1.33946	1.33945	1.33944	1.33943	1.33941	1.33940	1.33939	1.33938
28	1.33936	1.33935	1.33934	1.33933	1.33931	1.33930	1.33929	1.33928	1.33926	1.33925
29	1.33924	1.33922	1.33921	1.33920	1.33919	1.33917	1.33916	1.33915	1.33913	1.33912
30	1.33911	1.33909	1.33908	1.33907	1.33905	1.33904	1.33903	1.33901	1.33900	1.33899
31	1.33897	1.33896	1.33895	1.33893	1.33892	1.33891	1.33889	1.33888	1.33887	1.33885
32	1.33884	1.33882	1.33881	1.33880	1.33878	1.33877	1.33875	1.33874	1.33873	1.33871
33	1.33870	1.33868	1.33867	1.33865	1.33864	1.33863	1.33861	1.33860	1.33858	1.33857
34	1.33855	1.33854	1.33853	1.33851	1.33850	1.33848	1.33847	1.33845	1.33844	1.33842
35	1.33841	1.33839	1.33838	1.33836	1.33835	1.33833	1.33832	1.33830	1.33829	1.33827
36	1.33826	1.33824	1.33823	1.33821	1.33820	1.33818	1.33817	1.33815	1.33813	1.33812
37	1.33810	1.33809	1.33807	1.33806	1.33804	1.33802	1.33801	1.33799	1.33798	1.33796
38	1.33795	1.33793	1.33791	1.33790	1.33788	1.33787	1.33785	1.33783	1.33782	1.33780
39	1.33778	1.33777	1.33775	1.33774	1.33772	1.33770	1.33769	1.33767	1.33765	1.33764
40	1.33762									

AG5 Refractometer Calibration Fluid - R.I. Temperature Relationship



Code: 90-304 Issue: 1



Bellingham + Stanley Ltd.

Longfield Road, Tunbridge Wells

KENT. TN2 3EY

UK

Telephone: +44 (0) 1892 500400

Fax: +44 (0) 1892 543115

Bellingham + Stanley Inc.

1000 Hurricane Shoals Road

Bldg D, Suite 300

LAWRENCEVILLE, USA, GA 30043


Telephone: 770 822 6898

Fax: 770 822 9165

AG5 Refractometer Calibration Fluid °Brix Temperature Relationship

Temperature (°C)	0.0	0.1	0.2	0.3	0.4	0.5	0.6	0.7	0.8	0.9
10	5.58	5.58	5.57	5.57	5.56	5.56	5.55	5.55	5.54	5.54
11	5.53	5.53	5.52	5.52	5.51	5.51	5.50	5.50	5.49	5.49
12	5.48	5.48	5.47	5.46	5.46	5.45	5.45	5.44	5.44	5.43
13	5.43	5.42	5.42	5.41	5.40	5.40	5.39	5.39	5.38	5.38
14	5.37	5.37	5.36	5.35	5.35	5.34	5.34	5.33	5.33	5.32
15	5.31	5.31	5.30	5.30	5.29	5.28	5.28	5.27	5.27	5.26
16	5.25	5.25	5.24	5.24	5.23	5.22	5.22	5.21	5.21	5.20
17	5.19	5.19	5.18	5.17	5.17	5.16	5.16	5.15	5.14	5.14
18	5.13	5.12	5.12	5.11	5.10	5.10	5.09	5.08	5.08	5.07
19	5.07	5.06	5.05	5.05	5.04	5.03	5.03	5.02	5.01	5.01
20	5.00	4.99	4.98	4.98	4.97	4.96	4.96	4.95	4.94	4.94
21	4.93	4.92	4.92	4.91	4.90	4.89	4.89	4.88	4.87	4.87
22	4.86	4.85	4.84	4.84	4.83	4.82	4.82	4.81	4.80	4.79
23	4.79	4.78	4.77	4.76	4.76	4.75	4.74	4.73	4.73	4.72
24	4.71	4.70	4.70	4.69	4.68	4.67	4.66	4.66	4.65	4.64
25	4.63	4.63	4.62	4.61	4.60	4.59	4.59	4.58	4.57	4.56
26	4.55	4.55	4.54	4.53	4.52	4.51	4.51	4.50	4.49	4.48
27	4.47	4.47	4.46	4.45	4.44	4.43	4.42	4.42	4.41	4.40
28	4.39	4.38	4.37	4.36	4.36	4.35	4.34	4.33	4.32	4.31
29	4.30	4.30	4.29	4.28	4.27	4.26	4.25	4.24	4.23	4.23
30	4.22	4.21	4.20	4.19	4.18	4.17	4.16	4.15	4.14	4.14
31	4.13	4.12	4.11	4.10	4.09	4.08	4.07	4.06	4.05	4.04
32	4.03	4.02	4.01	4.01	4.00	3.99	3.98	3.97	3.96	3.95
33	3.94	3.93	3.92	3.91	3.90	3.89	3.88	3.87	3.86	3.85
34	3.84	3.83	3.82	3.81	3.80	3.79	3.78	3.77	3.76	3.75
35	3.74	3.73	3.72	3.71	3.70	3.69	3.68	3.67	3.66	3.65
36	3.64	3.63	3.62	3.61	3.60	3.59	3.58	3.57	3.56	3.54
37	3.53	3.52	3.51	3.50	3.49	3.48	3.47	3.46	3.45	3.44
38	3.43	3.42	3.41	3.39	3.38	3.37	3.36	3.35	3.34	3.33
39	3.32	3.31	3.29	3.28	3.27	3.26	3.25	3.24	3.23	3.22
40	3.20									

AG5 Refractometer Calibration Fluid - °Brix Temperature Relationship



Code: 90-304 Issue: 1



Bellingham + Stanley Ltd.

Longfield Road, Tunbridge Wells
KENT. TN2 3EY

UK
Telephone: +44 (0) 1892 500400
Fax: +44 (0) 1892 543115

Bellingham + Stanley Inc.

1000 Hurricane Shoals Road
Bldg D, Suite 300

LAWRENCEVILLE, USA, GA 30043
Telephone: 770 822 6898
Fax: 770 822 9165