


AG40 Refractometer Calibration Fluid Refractive Index Temperature Relationship

Temperature (°C)	0.0	0.1	0.2	0.3	0.4	0.5	0.6	0.7	0.8	0.9
10	1.40250	1.40248	1.40245	1.40242	1.40240	1.40237	1.40235	1.40232	1.40229	1.40227
11	1.40224	1.40222	1.40219	1.40217	1.40214	1.40211	1.40209	1.40206	1.40204	1.40201
12	1.40198	1.40196	1.40193	1.40191	1.40188	1.40185	1.40183	1.40180	1.40178	1.40175
13	1.40172	1.40170	1.40167	1.40164	1.40162	1.40159	1.40157	1.40154	1.40151	1.40149
14	1.40146	1.40143	1.40141	1.40138	1.40136	1.40133	1.40130	1.40128	1.40125	1.40122
15	1.40120	1.40117	1.40114	1.40112	1.40109	1.40106	1.40104	1.40101	1.40099	1.40096
16	1.40093	1.40091	1.40088	1.40085	1.40083	1.40080	1.40077	1.40075	1.40072	1.40069
17	1.40067	1.40064	1.40061	1.40059	1.40056	1.40053	1.40050	1.40048	1.40045	1.40042
18	1.40040	1.40037	1.40034	1.40032	1.40029	1.40026	1.40024	1.40021	1.40018	1.40015
19	1.40013	1.40010	1.40007	1.40005	1.40002	1.39999	1.39997	1.39994	1.39991	1.39988
20	1.39986	1.39983	1.39980	1.39977	1.39975	1.39972	1.39969	1.39967	1.39964	1.39961
21	1.39958	1.39956	1.39953	1.39950	1.39947	1.39945	1.39942	1.39939	1.39936	1.39934
22	1.39931	1.39928	1.39925	1.39923	1.39920	1.39917	1.39914	1.39911	1.39909	1.39906
23	1.39903	1.39900	1.39898	1.39895	1.39892	1.39889	1.39886	1.39884	1.39881	1.39878
24	1.39875	1.39873	1.39870	1.39867	1.39864	1.39861	1.39859	1.39856	1.39853	1.39850
25	1.39847	1.39844	1.39842	1.39839	1.39836	1.39833	1.39830	1.39828	1.39825	1.39822
26	1.39819	1.39816	1.39813	1.39811	1.39808	1.39805	1.39802	1.39799	1.39796	1.39794
27	1.39791	1.39788	1.39785	1.39782	1.39779	1.39776	1.39774	1.39771	1.39768	1.39765
28	1.39762	1.39759	1.39756	1.39753	1.39751	1.39748	1.39745	1.39742	1.39739	1.39736
29	1.39733	1.39730	1.39727	1.39725	1.39722	1.39719	1.39716	1.39713	1.39710	1.39707
30	1.39704	1.39701	1.39698	1.39695	1.39693	1.39690	1.39687	1.39684	1.39681	1.39678
31	1.39675	1.39672	1.39669	1.39666	1.39663	1.39660	1.39657	1.39654	1.39651	1.39648
32	1.39645	1.39643	1.39640	1.39637	1.39634	1.39631	1.39628	1.39625	1.39622	1.39619
33	1.39616	1.39613	1.39610	1.39607	1.39604	1.39601	1.39598	1.39595	1.39592	1.39589
34	1.39586	1.39583	1.39580	1.39577	1.39574	1.39571	1.39568	1.39565	1.39562	1.39559
35	1.39556	1.39553	1.39550	1.39547	1.39544	1.39541	1.39538	1.39534	1.39531	1.39528
36	1.39525	1.39522	1.39519	1.39516	1.39513	1.39510	1.39507	1.39504	1.39501	1.39498
37	1.39495	1.39492	1.39489	1.39485	1.39482	1.39479	1.39476	1.39473	1.39470	1.39467
38	1.39464	1.39461	1.39458	1.39455	1.39451	1.39448	1.39445	1.39442	1.39439	1.39436
39	1.39433	1.39430	1.39426	1.39423	1.39420	1.39417	1.39414	1.39411	1.39408	1.39404
40	1.39401									

AG40 Refractometer Calibration Fluid - R.I. Temperature Relationship



Code: 90-310 Issue: 1



Bellingham + Stanley Ltd.

Longfield Road, Tunbridge Wells

KENT. TN2 3EY

UK

Telephone: +44 (0) 1892 500400

Fax: +44 (0) 1892 543115

Bellingham + Stanley Inc.

1000 Hurricane Shoals Road

Bldg D, Suite 300

LAWRENCEVILLE, USA, GA 30043


Telephone: 770 822 6898

Fax: 770 822 9165

AG40 Refractometer Calibration Fluid °Brix Temperature Relationship

Temperature (°C)	0.0	0.1	0.2	0.3	0.4	0.5	0.6	0.7	0.8	0.9
10	41.35	41.34	41.33	41.31	41.30	41.29	41.27	41.26	41.25	41.23
11	41.22	41.21	41.19	41.18	41.17	41.16	41.14	41.13	41.12	41.10
12	41.09	41.08	41.06	41.05	41.04	41.02	41.01	41.00	40.98	40.97
13	40.96	40.94	40.93	40.92	40.90	40.89	40.88	40.86	40.85	40.84
14	40.82	40.81	40.80	40.78	40.77	40.76	40.74	40.73	40.71	40.70
15	40.69	40.67	40.66	40.65	40.63	40.62	40.61	40.59	40.58	40.57
16	40.55	40.54	40.52	40.51	40.50	40.48	40.47	40.46	40.44	40.43
17	40.42	40.40	40.39	40.37	40.36	40.35	40.33	40.32	40.31	40.29
18	40.28	40.26	40.25	40.24	40.22	40.21	40.20	40.18	40.17	40.15
19	40.14	40.13	40.11	40.10	40.08	40.07	40.06	40.04	40.03	40.01
20	40.00	39.99	39.97	39.96	39.94	39.93	39.92	39.90	39.89	39.87
21	39.86	39.85	39.83	39.82	39.80	39.79	39.77	39.76	39.75	39.73
22	39.72	39.70	39.69	39.68	39.66	39.65	39.63	39.62	39.60	39.59
23	39.58	39.56	39.55	39.53	39.52	39.50	39.49	39.48	39.46	39.45
24	39.43	39.42	39.40	39.39	39.37	39.36	39.35	39.33	39.32	39.30
25	39.29	39.27	39.26	39.24	39.23	39.21	39.20	39.19	39.17	39.16
26	39.14	39.13	39.11	39.10	39.08	39.07	39.05	39.04	39.02	39.01
27	38.99	38.98	38.96	38.95	38.93	38.92	38.91	38.89	38.88	38.86
28	38.85	38.83	38.82	38.80	38.79	38.77	38.76	38.74	38.73	38.71
29	38.70	38.68	38.67	38.65	38.64	38.62	38.61	38.59	38.58	38.56
30	38.55	38.53	38.51	38.50	38.48	38.47	38.45	38.44	38.42	38.41
31	38.39	38.38	38.36	38.35	38.33	38.32	38.30	38.29	38.27	38.26
32	38.24	38.22	38.21	38.19	38.18	38.16	38.15	38.13	38.12	38.10
33	38.08	38.07	38.05	38.04	38.02	38.01	37.99	37.98	37.96	37.94
34	37.93	37.91	37.90	37.88	37.87	37.85	37.83	37.82	37.80	37.79
35	37.77	37.76	37.74	37.72	37.71	37.69	37.68	37.66	37.64	37.63
36	37.61	37.60	37.58	37.56	37.55	37.53	37.52	37.50	37.48	37.47
37	37.45	37.44	37.42	37.40	37.39	37.37	37.35	37.34	37.32	37.31
38	37.29	37.27	37.26	37.24	37.22	37.21	37.19	37.18	37.16	37.14
39	37.13	37.11	37.09	37.08	37.06	37.04	37.03	37.01	36.99	36.98
40	36.96									

AG40 Refractometer Calibration Fluid - °Brix Temperature Relationship



Code: 90-310 Issue: 1



Bellingham + Stanley Ltd.

Longfield Road, Tunbridge Wells
 KENT. TN2 3EY
 UK
 Telephone: +44 (0) 1892 500400
 Fax: +44 (0) 1892 543115

Bellingham + Stanley Inc.

1000 Hurricane Shoals Road
 Bldg D, Suite 300
 LAWRENCEVILLE, USA, GA 30043
 Telephone: 770 822 6898
 Fax: 770 822 9165