


## AG10 Refractometer Calibration Fluid Refractive Index Temperature Relationship

Temperature (°C)	0.0	0.1	0.2	0.3	0.4	0.5	0.6	0.7	0.8	0.9
10	1.34886	1.34885	1.34884	1.34883	1.34882	1.34881	1.34880	1.34879	1.34878	1.34878
11	1.34877	1.34876	1.34875	1.34874	1.34873	1.34872	1.34871	1.34870	1.34869	1.34868
12	1.34867	1.34866	1.34865	1.34864	1.34863	1.34862	1.34861	1.34860	1.34859	1.34858
13	1.34857	1.34856	1.34855	1.34854	1.34853	1.34852	1.34851	1.34850	1.34849	1.34848
14	1.34847	1.34846	1.34845	1.34844	1.34843	1.34842	1.34841	1.34840	1.34839	1.34838
15	1.34837	1.34836	1.34835	1.34834	1.34833	1.34832	1.34831	1.34830	1.34829	1.34828
16	1.34827	1.34826	1.34825	1.34824	1.34823	1.34821	1.34820	1.34819	1.34818	1.34817
17	1.34816	1.34815	1.34814	1.34813	1.34812	1.34811	1.34810	1.34808	1.34807	1.34806
18	1.34805	1.34804	1.34803	1.34802	1.34801	1.34800	1.34798	1.34797	1.34796	1.34795
19	1.34794	1.34793	1.34792	1.34791	1.34789	1.34788	1.34787	1.34786	1.34785	1.34784
20	1.34782	1.34781	1.34780	1.34779	1.34778	1.34777	1.34775	1.34774	1.34773	1.34772
21	1.34771	1.34770	1.34768	1.34767	1.34766	1.34765	1.34764	1.34762	1.34761	1.34760
22	1.34759	1.34757	1.34756	1.34755	1.34754	1.34753	1.34751	1.34750	1.34749	1.34748
23	1.34746	1.34745	1.34744	1.34743	1.34741	1.34740	1.34739	1.34738	1.34736	1.34735
24	1.34734	1.34733	1.34731	1.34730	1.34729	1.34727	1.34726	1.34725	1.34724	1.34722
25	1.34721	1.34720	1.34718	1.34717	1.34716	1.34714	1.34713	1.34712	1.34710	1.34709
26	1.34708	1.34706	1.34705	1.34704	1.34702	1.34701	1.34700	1.34698	1.34697	1.34696
27	1.34694	1.34693	1.34692	1.34690	1.34689	1.34687	1.34686	1.34685	1.34683	1.34682
28	1.34681	1.34679	1.34678	1.34676	1.34675	1.34674	1.34672	1.34671	1.34669	1.34668
29	1.34666	1.34665	1.34664	1.34662	1.34661	1.34659	1.34658	1.34656	1.34655	1.34654
30	1.34652	1.34651	1.34649	1.34648	1.34646	1.34645	1.34643	1.34642	1.34640	1.34639
31	1.34637	1.34636	1.34634	1.34633	1.34631	1.34630	1.34628	1.34627	1.34625	1.34624
32	1.34622	1.34621	1.34619	1.34618	1.34616	1.34615	1.34613	1.34612	1.34610	1.34609
33	1.34607	1.34606	1.34604	1.34602	1.34601	1.34599	1.34598	1.34596	1.34595	1.34593
34	1.34591	1.34590	1.34588	1.34587	1.34585	1.34583	1.34582	1.34580	1.34579	1.34577
35	1.34575	1.34574	1.34572	1.34571	1.34569	1.34567	1.34566	1.34564	1.34562	1.34561
36	1.34559	1.34557	1.34556	1.34554	1.34553	1.34551	1.34549	1.34548	1.34546	1.34544
37	1.34542	1.34541	1.34539	1.34537	1.34536	1.34534	1.34532	1.34531	1.34529	1.34527
38	1.34525	1.34524	1.34522	1.34520	1.34519	1.34517	1.34515	1.34513	1.34512	1.34510
39	1.34508	1.34506	1.34505	1.34503	1.34501	1.34499	1.34498	1.34496	1.34494	1.34492
40	1.34490									

**AG10 Refractometer Calibration Fluid - R.I. Temperature Relationship**



Code: 90-306 Issue: 1



**Bellingham + Stanley Ltd.**

Longfield Road, Tunbridge Wells

KENT. TN2 3EY

UK

Telephone: +44 (0) 1892 500400

Fax: +44 (0) 1892 543115

**Bellingham + Stanley Inc.**

1000 Hurricane Shoals Road

Bldg D, Suite 300

LAWRENCEVILLE, USA, GA 30043


Telephone: 770 822 6898

Fax: 770 822 9165

## AG10 Refractometer Calibration Fluid °Brix Temperature Relationship

Temperature (°C)	0.0	0.1	0.2	0.3	0.4	0.5	0.6	0.7	0.8	0.9
10	10.67	10.66	10.65	10.65	10.64	10.64	10.63	10.62	10.62	10.61
11	10.61	10.60	10.59	10.59	10.58	10.58	10.57	10.56	10.56	10.55
12	10.54	10.54	10.53	10.53	10.52	10.51	10.51	10.50	10.49	10.49
13	10.48	10.48	10.47	10.46	10.46	10.45	10.44	10.44	10.43	10.42
14	10.42	10.41	10.40	10.40	10.39	10.38	10.38	10.37	10.36	10.36
15	10.35	10.34	10.34	10.33	10.32	10.32	10.31	10.30	10.30	10.29
16	10.28	10.28	10.27	10.26	10.26	10.25	10.24	10.24	10.23	10.22
17	10.21	10.21	10.20	10.19	10.19	10.18	10.17	10.17	10.16	10.15
18	10.14	10.14	10.13	10.12	10.11	10.11	10.10	10.09	10.09	10.08
19	10.07	10.06	10.06	10.05	10.04	10.03	10.03	10.02	10.01	10.00
20	10.00	9.99	9.98	9.97	9.97	9.96	9.95	9.94	9.94	9.93
21	9.92	9.91	9.90	9.90	9.89	9.88	9.87	9.87	9.86	9.85
22	9.84	9.83	9.83	9.82	9.81	9.80	9.79	9.79	9.78	9.77
23	9.76	9.75	9.75	9.74	9.73	9.72	9.71	9.71	9.70	9.69
24	9.68	9.67	9.66	9.66	9.65	9.64	9.63	9.62	9.61	9.61
25	9.60	9.59	9.58	9.57	9.56	9.55	9.55	9.54	9.53	9.52
26	9.51	9.50	9.49	9.48	9.48	9.47	9.46	9.45	9.44	9.43
27	9.42	9.41	9.41	9.40	9.39	9.38	9.37	9.36	9.35	9.34
28	9.33	9.32	9.32	9.31	9.30	9.29	9.28	9.27	9.26	9.25
29	9.24	9.23	9.22	9.21	9.20	9.20	9.19	9.18	9.17	9.16
30	9.15	9.14	9.13	9.12	9.11	9.10	9.09	9.08	9.07	9.06
31	9.05	9.04	9.03	9.02	9.01	9.00	8.99	8.98	8.97	8.96
32	8.95	8.94	8.93	8.92	8.91	8.90	8.89	8.88	8.87	8.86
33	8.85	8.84	8.83	8.82	8.81	8.80	8.79	8.78	8.77	8.76
34	8.75	8.74	8.73	8.72	8.71	8.70	8.69	8.68	8.67	8.66
35	8.65	8.64	8.63	8.61	8.60	8.59	8.58	8.57	8.56	8.55
36	8.54	8.53	8.52	8.51	8.50	8.49	8.47	8.46	8.45	8.44
37	8.43	8.42	8.41	8.40	8.39	8.37	8.36	8.35	8.34	8.33
38	8.32	8.31	8.30	8.28	8.27	8.26	8.25	8.24	8.23	8.22
39	8.20	8.19	8.18	8.17	8.16	8.15	8.13	8.12	8.11	8.10
40	8.09									

### AG10 Refractometer Calibration Fluid - °Brix Temperature Relationship



Code: 90-306 Issue: 1



#### **Bellingham + Stanley Ltd.**

Longfield Road, Tunbridge Wells  
 KENT. TN2 3EY  
 UK  
 Telephone: +44 (0) 1892 500400  
 Fax: +44 (0) 1892 543115

#### **Bellingham + Stanley Inc.**

1000 Hurricane Shoals Road  
 Bldg D, Suite 300  
 LAWRENCEVILLE, USA, GA 30043  
 Telephone: 770 822 6898  
 Fax: 770 822 9165