


AG10 Refractometer Calibration Fluid Refractive Index Temperature Relationship

Temperature (°C)	0.0	0.1	0.2	0.3	0.4	0.5	0.6	0.7	0.8	0.9
10	1.34886	1.34885	1.34884	1.34883	1.34882	1.34881	1.34880	1.34879	1.34878	1.34878
11	1.34877	1.34876	1.34875	1.34874	1.34873	1.34872	1.34871	1.34870	1.34869	1.34868
12	1.34867	1.34866	1.34865	1.34864	1.34863	1.34862	1.34861	1.34860	1.34859	1.34858
13	1.34857	1.34856	1.34855	1.34854	1.34853	1.34852	1.34851	1.34850	1.34849	1.34848
14	1.34847	1.34846	1.34845	1.34844	1.34843	1.34842	1.34841	1.34840	1.34839	1.34838
15	1.34837	1.34836	1.34835	1.34834	1.34833	1.34832	1.34831	1.34830	1.34829	1.34828
16	1.34827	1.34826	1.34825	1.34824	1.34823	1.34821	1.34820	1.34819	1.34818	1.34817
17	1.34816	1.34815	1.34814	1.34813	1.34812	1.34811	1.34810	1.34808	1.34807	1.34806
18	1.34805	1.34804	1.34803	1.34802	1.34801	1.34800	1.34798	1.34797	1.34796	1.34795
19	1.34794	1.34793	1.34792	1.34791	1.34789	1.34788	1.34787	1.34786	1.34785	1.34784
20	1.34782	1.34781	1.34780	1.34779	1.34778	1.34777	1.34775	1.34774	1.34773	1.34772
21	1.34771	1.34770	1.34768	1.34767	1.34766	1.34765	1.34764	1.34762	1.34761	1.34760
22	1.34759	1.34757	1.34756	1.34755	1.34754	1.34753	1.34751	1.34750	1.34749	1.34748
23	1.34746	1.34745	1.34744	1.34743	1.34741	1.34740	1.34739	1.34738	1.34736	1.34735
24	1.34734	1.34733	1.34731	1.34730	1.34729	1.34727	1.34726	1.34725	1.34724	1.34722
25	1.34721	1.34720	1.34718	1.34717	1.34716	1.34714	1.34713	1.34712	1.34710	1.34709
26	1.34708	1.34706	1.34705	1.34704	1.34702	1.34701	1.34700	1.34698	1.34697	1.34696
27	1.34694	1.34693	1.34692	1.34690	1.34689	1.34687	1.34686	1.34685	1.34683	1.34682
28	1.34681	1.34679	1.34678	1.34676	1.34675	1.34674	1.34672	1.34671	1.34669	1.34668
29	1.34666	1.34665	1.34664	1.34662	1.34661	1.34659	1.34658	1.34656	1.34655	1.34654
30	1.34652	1.34651	1.34649	1.34648	1.34646	1.34645	1.34643	1.34642	1.34640	1.34639
31	1.34637	1.34636	1.34634	1.34633	1.34631	1.34630	1.34628	1.34627	1.34625	1.34624
32	1.34622	1.34621	1.34619	1.34618	1.34616	1.34615	1.34613	1.34612	1.34610	1.34609
33	1.34607	1.34606	1.34604	1.34602	1.34601	1.34599	1.34598	1.34596	1.34595	1.34593
34	1.34591	1.34590	1.34588	1.34587	1.34585	1.34583	1.34582	1.34580	1.34579	1.34577
35	1.34575	1.34574	1.34572	1.34571	1.34569	1.34567	1.34566	1.34564	1.34562	1.34561
36	1.34559	1.34557	1.34556	1.34554	1.34553	1.34551	1.34549	1.34548	1.34546	1.34544
37	1.34542	1.34541	1.34539	1.34537	1.34536	1.34534	1.34532	1.34531	1.34529	1.34527
38	1.34525	1.34524	1.34522	1.34520	1.34519	1.34517	1.34515	1.34513	1.34512	1.34510
39	1.34508	1.34506	1.34505	1.34503	1.34501	1.34499	1.34498	1.34496	1.34494	1.34492
40	1.34490									

AG10 Refractometer Calibration Fluid - R.I. Temperature Relationship



Code: 90-306 Issue: 02 Revision: A



Bellingham + Stanley Ltd.

Longfield Road
Tunbridge Wells
Kent, TN2 3EY, UK
Telephone: +44 (0) 1892 500400
Fax: +44 (0) 1892 543115


Bellingham + Stanley Inc.

90 Horizon Drive
Suwanee
GA 30024, USA
Telephone: 678 804 5730
Fax: 678 804 5729

AG10 Refractometer Calibration Fluid °Brix Temperature Relationship

Temperature (°C)	0.0	0.1	0.2	0.3	0.4	0.5	0.6	0.7	0.8	0.9
10	10.67	10.66	10.65	10.65	10.64	10.64	10.63	10.62	10.62	10.61
11	10.61	10.60	10.59	10.59	10.58	10.58	10.57	10.56	10.56	10.55
12	10.54	10.54	10.53	10.53	10.52	10.51	10.51	10.50	10.49	10.49
13	10.48	10.48	10.47	10.46	10.46	10.45	10.44	10.44	10.43	10.42
14	10.42	10.41	10.40	10.40	10.39	10.38	10.38	10.37	10.36	10.36
15	10.35	10.34	10.34	10.33	10.32	10.32	10.31	10.30	10.30	10.29
16	10.28	10.28	10.27	10.26	10.26	10.25	10.24	10.24	10.23	10.22
17	10.21	10.21	10.20	10.19	10.19	10.18	10.17	10.17	10.16	10.15
18	10.14	10.14	10.13	10.12	10.11	10.11	10.10	10.09	10.09	10.08
19	10.07	10.06	10.06	10.05	10.04	10.03	10.03	10.02	10.01	10.00
20	10.00	9.99	9.98	9.97	9.97	9.96	9.95	9.94	9.94	9.93
21	9.92	9.91	9.90	9.90	9.89	9.88	9.87	9.87	9.86	9.85
22	9.84	9.83	9.83	9.82	9.81	9.80	9.79	9.79	9.78	9.77
23	9.76	9.75	9.75	9.74	9.73	9.72	9.71	9.71	9.70	9.69
24	9.68	9.67	9.66	9.66	9.65	9.64	9.63	9.62	9.61	9.61
25	9.60	9.59	9.58	9.57	9.56	9.55	9.55	9.54	9.53	9.52
26	9.51	9.50	9.49	9.48	9.48	9.47	9.46	9.45	9.44	9.43
27	9.42	9.41	9.41	9.40	9.39	9.38	9.37	9.36	9.35	9.34
28	9.33	9.32	9.32	9.31	9.30	9.29	9.28	9.27	9.26	9.25
29	9.24	9.23	9.22	9.21	9.20	9.20	9.19	9.18	9.17	9.16
30	9.15	9.14	9.13	9.12	9.11	9.10	9.09	9.08	9.07	9.06
31	9.05	9.04	9.03	9.02	9.01	9.00	8.99	8.98	8.97	8.96
32	8.95	8.94	8.93	8.92	8.91	8.90	8.89	8.88	8.87	8.86
33	8.85	8.84	8.83	8.82	8.81	8.80	8.79	8.78	8.77	8.76
34	8.75	8.74	8.73	8.72	8.71	8.70	8.69	8.68	8.67	8.66
35	8.65	8.64	8.63	8.61	8.60	8.59	8.58	8.57	8.56	8.55
36	8.54	8.53	8.52	8.51	8.50	8.49	8.47	8.46	8.45	8.44
37	8.43	8.42	8.41	8.40	8.39	8.37	8.36	8.35	8.34	8.33
38	8.32	8.31	8.30	8.28	8.27	8.26	8.25	8.24	8.23	8.22
39	8.20	8.19	8.18	8.17	8.16	8.15	8.13	8.12	8.11	8.10
40	8.09									

AG10 Refractometer Calibration Fluid - °Brix Temperature Relationship



Code: 90-306 Issue: 02 Revision: A



Bellingham + Stanley Ltd.
 Longfield Road
 Tunbridge Wells
 Kent, TN2 3EY, UK
 Telephone: +44 (0) 1892 500400
 Fax: +44 (0) 1892 543115

Bellingham + Stanley Inc.
 90 Horizon Drive
 Suwanee
 GA 30024, USA
 Telephone: 678 804 5730
 Fax: 678 804 5729