


## AG11.2 Refractometer Calibration Fluid Refractive Index Temperature Relationship

Temperature (°C)	0.0	0.1	0.2	0.3	0.4	0.5	0.6	0.7	0.8	0.9
10	1.35076	1.35075	1.35074	1.35073	1.35072	1.35071	1.35070	1.35069	1.35068	1.35067
11	1.35066	1.35065	1.35064	1.35063	1.35062	1.35061	1.35060	1.35059	1.35058	1.35057
12	1.35056	1.35055	1.35054	1.35053	1.35052	1.35051	1.35050	1.35049	1.35048	1.35047
13	1.35046	1.35045	1.35044	1.35043	1.35042	1.35041	1.35040	1.35039	1.35038	1.35037
14	1.35036	1.35035	1.35034	1.35033	1.35032	1.35031	1.35029	1.35028	1.35027	1.35026
15	1.35025	1.35024	1.35023	1.35022	1.35021	1.35020	1.35019	1.35018	1.35016	1.35015
16	1.35014	1.35013	1.35012	1.35011	1.35010	1.35009	1.35008	1.35007	1.35005	1.35004
17	1.35003	1.35002	1.35001	1.35000	1.34999	1.34998	1.34996	1.34995	1.34994	1.34993
18	1.34992	1.34991	1.34989	1.34988	1.34987	1.34986	1.34985	1.34984	1.34982	1.34981
19	1.34980	1.34979	1.34978	1.34977	1.34975	1.34974	1.34973	1.34972	1.34971	1.34969
20	1.34968	1.34967	1.34966	1.34965	1.34963	1.34962	1.34961	1.34960	1.34959	1.34957
21	1.34956	1.34955	1.34954	1.34952	1.34951	1.34950	1.34949	1.34947	1.34946	1.34945
22	1.34944	1.34942	1.34941	1.34940	1.34939	1.34937	1.34936	1.34935	1.34934	1.34932
23	1.34931	1.34930	1.34928	1.34927	1.34926	1.34924	1.34923	1.34922	1.34921	1.34919
24	1.34918	1.34917	1.34915	1.34914	1.34913	1.34911	1.34910	1.34909	1.34907	1.34906
25	1.34905	1.34903	1.34902	1.34901	1.34899	1.34898	1.34897	1.34895	1.34894	1.34893
26	1.34891	1.34890	1.34888	1.34887	1.34886	1.34884	1.34883	1.34881	1.34880	1.34879
27	1.34877	1.34876	1.34874	1.34873	1.34872	1.34870	1.34869	1.34867	1.34866	1.34865
28	1.34863	1.34862	1.34860	1.34859	1.34857	1.34856	1.34855	1.34853	1.34852	1.34850
29	1.34849	1.34847	1.34846	1.34844	1.34843	1.34841	1.34840	1.34838	1.34837	1.34835
30	1.34834	1.34832	1.34831	1.34829	1.34828	1.34826	1.34825	1.34823	1.34822	1.34820
31	1.34819	1.34817	1.34816	1.34814	1.34813	1.34811	1.34810	1.34808	1.34807	1.34805
32	1.34804	1.34802	1.34800	1.34799	1.34797	1.34796	1.34794	1.34793	1.34791	1.34789
33	1.34788	1.34786	1.34785	1.34783	1.34781	1.34780	1.34778	1.34777	1.34775	1.34773
34	1.34772	1.34770	1.34769	1.34767	1.34765	1.34764	1.34762	1.34760	1.34759	1.34757
35	1.34755	1.34754	1.34752	1.34750	1.34749	1.34747	1.34745	1.34744	1.34742	1.34740
36	1.34739	1.34737	1.34735	1.34734	1.34732	1.34730	1.34729	1.34727	1.34725	1.34723
37	1.34722	1.34720	1.34718	1.34717	1.34715	1.34713	1.34711	1.34710	1.34708	1.34706
38	1.34704	1.34703	1.34701	1.34699	1.34697	1.34696	1.34694	1.34692	1.34690	1.34689
39	1.34687	1.34685	1.34683	1.34681	1.34680	1.34678	1.34676	1.34674	1.34672	1.34670
40	1.34669									

**AG11.2 Refractometer Calibration Fluid - R.I. Temperature Relationship**



Code: 90-307 Issue: 02 Revision: A



**Bellingham + Stanley Ltd.**

Longfield Road  
Tunbridge Wells  
Kent, TN2 3EY, UK  
Telephone: +44 (0) 1892 500400  
Fax: +44 (0) 1892 543115


**Bellingham + Stanley Inc.**

90 Horizon Drive  
Suwanee  
GA 30024, USA  
Telephone: 678 804 5730  
Fax: 678 804 5729

## AG11.2 Refractometer Calibration Fluid °Brix Temperature Relationship

Temperature (°C)	0.0	0.1	0.2	0.3	0.4	0.5	0.6	0.7	0.8	0.9
10	11.89	11.88	11.88	11.87	11.86	11.86	11.85	11.84	11.84	11.83
11	11.83	11.82	11.81	11.81	11.80	11.79	11.79	11.78	11.77	11.77
12	11.76	11.76	11.75	11.74	11.74	11.73	11.72	11.72	11.71	11.70
13	11.70	11.69	11.68	11.68	11.67	11.66	11.66	11.65	11.64	11.64
14	11.63	11.62	11.62	11.61	11.60	11.60	11.59	11.58	11.58	11.57
15	11.56	11.56	11.55	11.54	11.53	11.53	11.52	11.51	11.51	11.50
16	11.49	11.49	11.48	11.47	11.46	11.46	11.45	11.44	11.44	11.43
17	11.42	11.41	11.41	11.40	11.39	11.38	11.38	11.37	11.36	11.36
18	11.35	11.34	11.33	11.33	11.32	11.31	11.30	11.30	11.29	11.28
19	11.27	11.27	11.26	11.25	11.24	11.24	11.23	11.22	11.21	11.20
20	11.20	11.19	11.18	11.17	11.17	11.16	11.15	11.14	11.13	11.13
21	11.12	11.11	11.10	11.09	11.09	11.08	11.07	11.06	11.05	11.05
22	11.04	11.03	11.02	11.01	11.01	11.00	10.99	10.98	10.97	10.96
23	10.96	10.95	10.94	10.93	10.92	10.92	10.91	10.90	10.89	10.88
24	10.87	10.86	10.86	10.85	10.84	10.83	10.82	10.81	10.80	10.80
25	10.79	10.78	10.77	10.76	10.75	10.74	10.74	10.73	10.72	10.71
26	10.70	10.69	10.68	10.67	10.66	10.66	10.65	10.64	10.63	10.62
27	10.61	10.60	10.59	10.58	10.57	10.57	10.56	10.55	10.54	10.53
28	10.52	10.51	10.50	10.49	10.48	10.47	10.46	10.45	10.44	10.44
29	10.43	10.42	10.41	10.40	10.39	10.38	10.37	10.36	10.35	10.34
30	10.33	10.32	10.31	10.30	10.29	10.28	10.27	10.26	10.25	10.24
31	10.23	10.22	10.21	10.20	10.19	10.18	10.17	10.16	10.15	10.14
32	10.13	10.12	10.11	10.10	10.09	10.08	10.07	10.06	10.05	10.04
33	10.03	10.02	10.01	10.00	9.99	9.98	9.97	9.96	9.95	9.94
34	9.93	9.92	9.91	9.90	9.89	9.87	9.86	9.85	9.84	9.83
35	9.82	9.81	9.80	9.79	9.78	9.77	9.76	9.75	9.73	9.72
36	9.71	9.70	9.69	9.68	9.67	9.66	9.65	9.64	9.62	9.61
37	9.60	9.59	9.58	9.57	9.56	9.55	9.53	9.52	9.51	9.50
38	9.49	9.48	9.47	9.45	9.44	9.43	9.42	9.41	9.40	9.39
39	9.37	9.36	9.35	9.34	9.33	9.32	9.30	9.29	9.28	9.27
40	9.26									

**AG11.2 Refractometer Calibration Fluid - °Brix Temperature Relationship**



Code: 90-307 Issue: 02 Revision: A



**Bellingham + Stanley Ltd.**

Longfield Road  
Tunbridge Wells  
Kent, TN2 3EY, UK  
Telephone: +44 (0) 1892 500400  
Fax: +44 (0) 1892 543115

**Bellingham + Stanley Inc.**

90 Horizon Drive  
Suwanee  
GA 30024, USA  
Telephone: 678 804 5730  
Fax: 678 804 5729