


AG12 Refractometer Calibration Fluid Refractive Index Temperature Relationship

Temperature (°C)	0.0	0.1	0.2	0.3	0.4	0.5	0.6	0.7	0.8	0.9
10	1.35204	1.35203	1.35202	1.35201	1.35200	1.35199	1.35198	1.35197	1.35196	1.35195
11	1.35194	1.35193	1.35192	1.35191	1.35190	1.35189	1.35188	1.35187	1.35186	1.35185
12	1.35184	1.35183	1.35181	1.35180	1.35179	1.35178	1.35177	1.35176	1.35175	1.35174
13	1.35173	1.35172	1.35171	1.35170	1.35169	1.35168	1.35167	1.35166	1.35165	1.35163
14	1.35162	1.35161	1.35160	1.35159	1.35158	1.35157	1.35156	1.35155	1.35154	1.35153
15	1.35151	1.35150	1.35149	1.35148	1.35147	1.35146	1.35145	1.35144	1.35143	1.35141
16	1.35140	1.35139	1.35138	1.35137	1.35136	1.35135	1.35133	1.35132	1.35131	1.35130
17	1.35129	1.35128	1.35127	1.35125	1.35124	1.35123	1.35122	1.35121	1.35120	1.35118
18	1.35117	1.35116	1.35115	1.35114	1.35112	1.35111	1.35110	1.35109	1.35108	1.35106
19	1.35105	1.35104	1.35103	1.35102	1.35100	1.35099	1.35098	1.35097	1.35096	1.35094
20	1.35093	1.35092	1.35091	1.35089	1.35088	1.35087	1.35086	1.35084	1.35083	1.35082
21	1.35081	1.35079	1.35078	1.35077	1.35076	1.35074	1.35073	1.35072	1.35070	1.35069
22	1.35068	1.35067	1.35065	1.35064	1.35063	1.35061	1.35060	1.35059	1.35058	1.35056
23	1.35055	1.35054	1.35052	1.35051	1.35050	1.35048	1.35047	1.35046	1.35044	1.35043
24	1.35042	1.35040	1.35039	1.35038	1.35036	1.35035	1.35034	1.35032	1.35031	1.35029
25	1.35028	1.35027	1.35025	1.35024	1.35023	1.35021	1.35020	1.35018	1.35017	1.35016
26	1.35014	1.35013	1.35011	1.35010	1.35009	1.35007	1.35006	1.35004	1.35003	1.35002
27	1.35000	1.34999	1.34997	1.34996	1.34994	1.34993	1.34992	1.34990	1.34989	1.34987
28	1.34986	1.34984	1.34983	1.34981	1.34980	1.34978	1.34977	1.34975	1.34974	1.34973
29	1.34971	1.34970	1.34968	1.34967	1.34965	1.34964	1.34962	1.34961	1.34959	1.34958
30	1.34956	1.34955	1.34953	1.34951	1.34950	1.34948	1.34947	1.34945	1.34944	1.34942
31	1.34941	1.34939	1.34938	1.34936	1.34934	1.34933	1.34931	1.34930	1.34928	1.34927
32	1.34925	1.34924	1.34922	1.34920	1.34919	1.34917	1.34916	1.34914	1.34912	1.34911
33	1.34909	1.34908	1.34906	1.34904	1.34903	1.34901	1.34899	1.34898	1.34896	1.34895
34	1.34893	1.34891	1.34890	1.34888	1.34886	1.34885	1.34883	1.34881	1.34880	1.34878
35	1.34876	1.34875	1.34873	1.34871	1.34870	1.34868	1.34866	1.34864	1.34863	1.34861
36	1.34859	1.34858	1.34856	1.34854	1.34852	1.34851	1.34849	1.34847	1.34846	1.34844
37	1.34842	1.34840	1.34839	1.34837	1.34835	1.34833	1.34832	1.34830	1.34828	1.34826
38	1.34825	1.34823	1.34821	1.34819	1.34817	1.34816	1.34814	1.34812	1.34810	1.34808
39	1.34807	1.34805	1.34803	1.34801	1.34799	1.34797	1.34796	1.34794	1.34792	1.34790
40	1.34788									

AG12 Refractometer Calibration Fluid - R.I. Temperature Relationship



Code: 90-311 Issue: 02 Revision: A




Bellingham + Stanley Ltd.
 Longfield Road
 Tunbridge Wells
 Kent, TN2 3EY, UK
 Telephone: +44 (0) 1892 500400
 Fax: +44 (0) 1892 543115

Bellingham + Stanley Inc.
 90 Horizon Drive
 Suwanee
 GA 30024, USA
 Telephone: 678 804 5730
 Fax: 678 804 5729

AG12 Refractometer Calibration Fluid °Brix Temperature Relationship

Temperature (°C)	0.0	0.1	0.2	0.3	0.4	0.5	0.6	0.7	0.8	0.9
10	12.70	12.70	12.69	12.68	12.68	12.67	12.66	12.66	12.65	12.65
11	12.64	12.63	12.63	12.62	12.61	12.61	12.60	12.59	12.59	12.58
12	12.57	12.57	12.56	12.55	12.55	12.54	12.53	12.53	12.52	12.51
13	12.51	12.50	12.49	12.49	12.48	12.47	12.47	12.46	12.45	12.45
14	12.44	12.43	12.43	12.42	12.41	12.40	12.40	12.39	12.38	12.38
15	12.37	12.36	12.36	12.35	12.34	12.33	12.33	12.32	12.31	12.31
16	12.30	12.29	12.28	12.28	12.27	12.26	12.25	12.25	12.24	12.23
17	12.23	12.22	12.21	12.20	12.20	12.19	12.18	12.17	12.17	12.16
18	12.15	12.14	12.14	12.13	12.12	12.11	12.11	12.10	12.09	12.08
19	12.07	12.07	12.06	12.05	12.04	12.04	12.03	12.02	12.01	12.00
20	12.00	11.99	11.98	11.97	11.97	11.96	11.95	11.94	11.93	11.93
21	11.92	11.91	11.90	11.89	11.88	11.88	11.87	11.86	11.85	11.84
22	11.84	11.83	11.82	11.81	11.80	11.79	11.79	11.78	11.77	11.76
23	11.75	11.74	11.74	11.73	11.72	11.71	11.70	11.69	11.68	11.68
24	11.67	11.66	11.65	11.64	11.63	11.62	11.62	11.61	11.60	11.59
25	11.58	11.57	11.56	11.55	11.55	11.54	11.53	11.52	11.51	11.50
26	11.49	11.48	11.47	11.47	11.46	11.45	11.44	11.43	11.42	11.41
27	11.40	11.39	11.38	11.37	11.36	11.36	11.35	11.34	11.33	11.32
28	11.31	11.30	11.29	11.28	11.27	11.26	11.25	11.24	11.23	11.22
29	11.21	11.21	11.20	11.19	11.18	11.17	11.16	11.15	11.14	11.13
30	11.12	11.11	11.10	11.09	11.08	11.07	11.06	11.05	11.04	11.03
31	11.02	11.01	11.00	10.99	10.98	10.97	10.96	10.95	10.94	10.93
32	10.92	10.91	10.90	10.89	10.88	10.87	10.86	10.85	10.84	10.83
33	10.82	10.81	10.80	10.78	10.77	10.76	10.75	10.74	10.73	10.72
34	10.71	10.70	10.69	10.68	10.67	10.66	10.65	10.64	10.63	10.61
35	10.60	10.59	10.58	10.57	10.56	10.55	10.54	10.53	10.52	10.51
36	10.49	10.48	10.47	10.46	10.45	10.44	10.43	10.42	10.41	10.39
37	10.38	10.37	10.36	10.35	10.34	10.33	10.31	10.30	10.29	10.28
38	10.27	10.26	10.25	10.23	10.22	10.21	10.20	10.19	10.18	10.16
39	10.15	10.14	10.13	10.12	10.11	10.09	10.08	10.07	10.06	10.05
40	10.03									

AG12 Refractometer Calibration Fluid - °Brix Temperature Relationship



Code: 90-311 Issue: 02 Revision: A



Bellingham + Stanley Ltd.
 Longfield Road
 Tunbridge Wells
 Kent, TN2 3EY, UK
 Telephone: +44 (0) 1892 500400
 Fax: +44 (0) 1892 543115

Bellingham + Stanley Inc.
 90 Horizon Drive
 Suwanee
 GA 30024, USA
 Telephone: 678 804 5730
 Fax: 678 804 5729